

Maintenance Allocation Chart

6515-01-291-1199

Defibrillator ECG Monitor/Recorder, Model HP 43110MC

[C – Operator or Crew; O – Unit Maintenance; DS – Direct Support Maintenance; GS – General Support Maintenance; D – Depot Maintenance]

(1) GROUP NUMBER	(2) ASSEMBLY GROUP	(3) MAINTENANCE FUNCTION	(4) MAINTENANCE LEVEL	(5) TOOLS AND EQUIPMENT	(6) REMARKS
01	Monitor/Recorder Module	Inspect	C		Tools and Equipment not required for crew inspection.
		Service	C		Tools and Equipment not required for crew servicing.
		Test	C		Tools and Equipment not required for crew testing.
		Calibrate	O	01, 12, 20, 29	
		Repair	O	01, 12, 20, 29	
		Safety	O	01, 12, 20, 29	
02	Defibrillator Module	Inspect	C		Tools and Equipment not required for crew inspection.
		Service	C		Tools and Equipment not required for crew servicing.
		Test	C		Tools and Equipment not required for crew testing.
		Calibrate	O	01, 12, 20, 27	
		Repair	O	01, 12, 20, 27	
		Safety	O	01, 12, 20, 27	
03	Monitor/Recorder Module Battery	Test	C		Tools and Equipment not required for crew testing.
		Service	O	01	
		Inspect	O	01, 12	
		Replace	O		Tools and Equipment not required for replacing.
04	Defibrillator Module Battery	Test	C		Tools and Equipment not required for crew testing.
		Service	O	01	
		Inspect	O	01, 12	
		Replace	O		Tools and Equipment not required for replacing.